

P001 樹林焚化廠CEMS系統日報表-A

報表日期: 111年11月01日

項目 單位 小時\限標	#1-HCl_C	#1-CO_C	#1-NOx_C	#1-SO2_C	#1-H2O	#1-O2_C	#1-OPAC	#1-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	6.54	0.72	55.51	0.58	19.19	11.19	1.47	77814.08
1	9.75	1.41	61.14	0.46	19.95	11.13	1.46	77618.25
2	5.61	1.11	62.32	0.39	20.41	10.91	1.46	78807.66
3	7.49	6.30	51.29	0.48	19.89	10.78	1.47	80143.74
4	7.31	2.25	49.17	0.46	20.35	10.87	1.47	79231.91
5	10.91	4.15	50.67	0.15	20.46	10.79	1.45	82128.93
6	8.64	3.33	51.36	0.10	21.24	10.81	1.45	78935.23
7	5.30	4.43	66.74	0.08	19.46	11.90	1.47	72022.75
8	0.80	8.34	64.42	0.13	20.80	11.86	1.68	70862.05
9	1.40	0.51	42.22	0.19	23.14	10.53	1.47	79166.94
10	7.81	2.51	43.85	0.15	21.86	10.17	1.44	83293.98
11	9.50	5.59	44.59	0.14	20.42	10.69	1.45	80868.67
12	7.98	8.59	43.74	0.09	19.87	11.00	1.46	77989.63
13	10.19	3.14	43.18	0.31	18.34	11.02	1.45	77519.44
14	9.70	1.54	48.57	0.50	17.81	11.30	1.46	75613.15
15	9.45	1.59	53.50	0.35	17.96	11.23	1.46	77927.08
16	5.61	2.53	54.75	0.28	17.47	11.90	1.46	73195.05
17	7.56	1.56	57.16	0.73	18.42	11.36	1.45	77000.90
18	7.49	1.00	64.72	0.19	20.41	10.72	1.46	80999.19
19	10.19	1.07	63.10	0.15	19.48	10.85	1.46	80973.56
20	9.28	1.75	57.56	0.15	18.82	11.28	1.46	78086.99
21	11.35	1.26	59.39	0.22	18.92	11.09	1.46	79516.10
22	10.91	1.11	58.34	0.12	19.65	10.68	1.45	82073.14
23	6.15	1.14	59.90	0.09	18.79	11.58	1.45	74055.98
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	7.79	2.79	54.47	0.27	19.71	11.07	1.47	78160.18
最 大 值	11.35	8.59	66.74	0.73	23.14	11.90	2.39	83293.98
發生時間	21	12	7	17	9	7		10
最 小 值	0.80	0.51	42.22	0.08	17.47	10.17	1.41	70862.05
發生時間	8	9	9	7	16	10		8
零點偏移(率)	-0.44 %	0.00 %	-0.23 %	-0.03 %		0.05	0.00	0.00 %
全幅偏移(率)	1.34 %	-0.27 %	-0.36 %	0.12 %		0.06	0.00	0.30 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P002 樹林焚化廠CEMS系統日報表-A

報表日期: 111年11月01日

項目 單位 小時\限標	#2-HCl_C	#2-CO_C	#2-NOx_C	#2-SO2_C	#2-H2O	#2-O2_C	#2-OPAC	#2-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	9.72	2.98	50.30	0.81	15.57	11.49	1.25	87696.16
1	9.08	2.72	67.12	0.14	18.72	11.54	1.25	84050.74
2	10.75	4.07	52.57	0.85	18.10	10.70	1.25	92334.50
3	10.73	1.29	63.00	1.06	18.30	10.71	1.25	91626.18
4	11.15	2.73	58.12	0.81	16.98	11.25	1.25	88161.10
5	9.95	6.27	51.70	0.69	17.09	11.48	1.24	85665.88
6	7.63	3.23	51.82	0.40	17.82	10.99	1.30	95099.22
7	9.34	6.91	60.40	0.59	16.54	11.67	1.29	85077.99
8	10.23	2.57	65.10	0.41	18.56	10.86	1.28	90217.96
9	9.67	17.36	48.25	0.84	16.12	11.22	1.27	86900.53
10	9.36	15.49	56.15	1.06	14.48	11.87	1.29	82671.87
11	10.46	1.53	57.58	0.68	16.23	11.56	1.27	84453.47
12	10.11	2.31	52.02	0.68	16.23	11.30	1.28	86013.04
13	10.35	3.24	57.81	1.01	14.53	11.70	1.29	85096.45
14	12.10	13.94	55.53	2.09	14.12	11.82	1.27	80570.53
15	17.48	13.61	55.57	2.84	15.11	11.62	1.26	84168.62
16	8.56	7.09	61.07	2.47	14.03	11.92	1.32	83199.39
17	10.37	1.09	56.07	1.01	14.12	11.97	1.33	82312.58
18	10.51	0.86	66.82	1.12	14.64	11.94	1.32	84664.75
19	10.93	0.74	59.30	0.41	17.18	11.13	1.32	88909.46
20	10.58	1.73	56.16	0.42	17.50	10.99	1.33	89967.11
21	10.91	1.31	55.54	0.61	16.18	11.33	1.33	87565.98
22	9.38	1.33	65.06	0.47	16.34	11.82	1.33	83111.12
23	10.56	1.36	62.46	0.28	16.92	11.55	1.33	86056.93
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	10.41	4.82	57.73	0.91	16.31	11.43	1.29	86482.98
最 大 值	17.48	17.36	67.12	2.84	18.72	11.97	1.34	95099.22
發生時間	15	9	1	15	1	17		6
最 小 值	7.63	0.74	48.25	0.14	14.03	10.70	1.23	80570.53
發生時間	6	19	9	1	16	2		14
零點偏移(率)	-0.05 %	-0.00 %	-0.23 %	-0.03 %		-0.06	0.00	0.00 %
全幅偏移(率)	0.03 %	-0.06 %	-0.14 %	0.03 %		-0.03	-0.03	0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								

P003 樹林焚化廠CEMS系統日報表-A

報表日期: 111年11月01日

項目 單位 小時\限標	#3-HCl_C	#3-CO_C	#3-NOx_C	#3-SO2_C	#3-H2O	#3-O2_C	#3-OPAC	#3-FLOW_C
	ppm	ppm	ppm	ppm	%	%	%	m3/Hour
	60.00	150.00	220.00	150.00	100.00	25.00	10.00	999999.00
0	8.49	9.21	59.61	3.02	14.41	11.40	1.29	74155.96
1	7.82	1.66	65.71	2.44	16.75	11.65	1.30	72361.50
2	9.70	1.22	59.72	2.94	15.73	10.99	1.30	77690.15
3	8.36	1.37	60.03	2.54	16.60	11.01	1.31	78308.60
4	9.12	2.35	60.56	2.84	16.19	11.03	1.31	78545.71
5	10.68	3.79	62.93	3.14	15.45	11.18	1.33	77117.93
6	9.13	4.14	58.54	2.94	16.64	10.85	1.30	79080.64
7	1.10	3.20	53.27	1.62	16.25	10.96	1.30	75564.13
8	3.19	3.43	63.39	1.76	16.48	11.37	1.30	77245.67
9	7.58	24.72	55.85	2.21	16.96	10.71	1.30	80786.37
10	6.30	7.79	46.11	2.36	16.43	10.37	1.30	78267.53
11	3.89	1.81	47.15	1.79	16.02	10.67	1.30	75797.98
12	4.04	3.65	51.84	2.07	16.00	10.79	1.30	76413.17
13	8.15	4.20	52.22	2.26	15.94	10.80	1.30	77315.03
14	9.99	3.64	55.71	2.32	14.84	11.19	1.31	77364.48
15	14.71	12.78	56.94	3.26	14.24	11.32	1.29	74899.66
16	13.27	6.44	51.46	2.84	13.98	11.07	1.26	71959.67
17	14.66	2.39	60.46	2.93	14.12	11.22	1.27	75553.56
18	7.06	4.68	61.28	3.38	13.19	11.80	1.28	71142.23
19	12.04	1.58	60.81	2.41	15.82	11.05	1.28	74786.00
20	7.53	1.45	56.16	1.75	17.24	10.54	1.29	78531.12
21	8.50	1.00	58.36	1.83	16.45	10.90	1.28	76797.98
22	6.97	1.46	57.38	1.75	17.30	10.67	1.28	77574.30
23	8.80	1.91	55.59	1.84	16.84	10.76	1.28	77592.22
監測時數	24	24	24	24	24	24	24	24
超限次數	0	0	0	0	0	0	0	0
有效時數	24	24	24	24	24	24	24	24
平 均 值	8.38	4.58	57.13	2.43	15.83	11.01	1.29	76452.15
最 大 值	14.71	24.72	65.71	3.38	17.30	11.80	1.33	80786.37
發生時間	15	9	1	18	22	18		9
最 小 值	1.10	1.00	46.11	1.62	13.19	10.37	1.25	71142.23
發生時間	7	21	10	7	18	10		18
零點偏移(率)	-0.07 %	-0.00 %	-0.07 %	-0.01 %		0.02	0.00	0.00 %
全幅偏移(率)	-0.32 %	0.03 %	-0.21 %	-0.01 %		0.23	-0.03	0.20 %
有效比率	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %	100.00 %
備註								